

AT-Nr. / AT-No.	AT-Nr. Selbsttest / AT-No. self-test	Ref-Nr / ID-No. *	Hersteller / Manufacturer	Testname / Test name	Zielantigen / target antigen	Sensitivität / Sensitivity			Gesamt-Sensitivität / total sensitivity
						Cq ≤25	Cq 25-30	Cq ≥30	
AT255/20		C8602C	Nanjing Vazyme Medical Technology Co., Ltd.	Vazyme SARS-CoV-2 Antigen Detection Kit (Colloidal Gold-Based)	N+S	85,0%	5,0%	0,0%	36,0%
AT090/20		FRC0G 020P	NanoEntek Inc	FREND™ COVID-19 Ag	N	88,2%	8,7%	0,0%	34,0%
AT517/20		17252668	Nantong Diagnos Biotechnology Co., Ltd.	COVID-19 Antigen Saliva Test Kit (Colloidal Gold)	N	100,0%	56,5%	0,0%	60,0%
AT425/21		n.a.	Nantong Egens Biotechnology	SARS-CoV-2 Antigen Rapid Test	N	100,0%	100,0%	50,0%	90,0%
AT722/21		SVRAG	Neo-nostics (Suzhou) Bioengineering Co., Ltd.	COVID 19 Antigen Test Kit (Colloidal Gold Method)	N	100,0%	75,0%	10,0%	72,0%
AT331/21	AT1210/21	COVID-19-NG08	New Gene (Hangzhou) Bioengineering Co., Ltd.	Covid-19-Antigen-Testkit	N	100,0%	87,0%	20,0%	78,0%
AT989/21		NBMLRS-XG-S2	Ningbo Beautiful Life Medical Biotechnology	Saliva SARS-CoV-2(2019-nCoV) Antigen Test Kit (Nanocarbon Assay)	N	100,0%	35,0%	0,0%	54,0%
AT036/21		MY28	Novatech Tibbi Cihaz Ürünleri San. Ve Tic.A.S.	novacheck®-Ag SARS-CoV-2 Covid-19 Antigen Rapid Test	N	94,1%	21,7%	0,0%	42,0%
AT117/20		CAT-01	Oncosem Onkolojik Sistemler San. Ve Tic.A.S.	CAT Antigen Covid Rapid Test	N	94,1%	30,4%	0,0%	46,0%
AT538/20		COV04S	PCL, Inc.	PCL COVID19 Ag Gold Saliva	N	100,0%	52,2%	0,0%	58,0%
AT209/20		n.a.	PerGrande BioTech Development Co., Ltd.	SARS-CoV-2 Antigen Detection Kit (Colloidal Gold Immunochromatographic Assay)	N+S	100,0%	17,4%	0,0%	42,0%
AT078/20		PR-FC13	Precision Biosensor Inc. (Axon Lab AG)	Exdia COVID-19-Ag-Test	N	100,0%	60,9%	0,0%	64,0%
AT376/20		V1330	ProGnosis Biotech	Rapid Test Ag 2019-nCoV	N	94,1%	65,2%	10,0%	64,0%
AT863/21		n.a.	Pro-med (Beijing) Technology Co.,Ltd.	Pro-med COVID-19 Antigen Rapid Detection Kit	N	80,0%	5,0%	0,0%	34,0%
AT935/21		n.a.	Qingdao AIBO Diagnostic CO., Ltd.	Novel Coronavirus (COVID-19) Antigen Rapid Test Kit	N	100,0%	20,0%	0,0%	48,0%
AT320/20	AT1224/21	H100G	Qingdao Hightop Biotech Co., Ltd.	Hightop SARS-CoV-2 (Covid-19) Antigen Rapid Test	N	100,0%	43,5%	0,0%	54,0%
AT008/20		N6803	R-Biopharm AG	RIDA®QUICK SARS-CoV-2 Antigen	N	100,0%	17,4%	0,0%	44,0%
AT199/20		COV Ag-6012	Safecare Biotech Hangzhou Co., Ltd.	Safecare COVID-19 Ag Rapid Test Kit (Swab)	N	100,0%	60,9%	0,0%	62,0%
AT106/20		COV-06-S	Salofa OY	salocor SARS-CoV-2 Antigen Rapid Test Cassette (Nasopharyngeal swab)	N	82,4%	13,0%	0,0%	34,0%
AT506/21		S3109E	Sansure Biotech Inc.	SARS-CoV-2 Rapid Antigen Test (Colloidal Gold Method)	N	100,0%	25,0%	0,0%	50,0%
AT069/20		34	ScheBo Biotech AG	ScheBo SARS-CoV-2 Quick Antigen	N	100,0%	91,3%	10,0%	78,0%
AT023/20 (AT004/20)	AT1154/21	Q-NCOV-01G (9901-NCOV-01G)	SD Biosensor, Inc. (Roche Diagnostics GmbH)	STANDARD Q COVID-19 Ag Test (SARS-CoV-2 Rapid Antigen Test)	N	88,9%	30,4%	0,0%	46,0%